



# PRODUCT CONFORMITY CERTIFICATE

This is certify that the

## **OXITEC 5000/ SME 5 Oxygen Analyser**

manufactured by:

### **ENOTEC GmbH**

*Höher Birken 6  
51709 Marienheide  
Germany*

has been assessed by Sira Certification Service  
and found to comply with:

### **MCERTS Performance Standards for Continuous Emission Monitoring Systems, Version 2, Revision 1 (April 2003)**

Certification Ranges :

O <sub>2</sub>	0 to 6 % vol
O <sub>2</sub>	0 to 12 % vol
O <sub>2</sub>	0 to 25 % vol

*Certification is awarded in respect of the conditions stated in this certificate*

Project No: 674/0159  
Certificate No: Sira MC 050063/00  
Initial Certification: 06 October 2005  
This Certificate Issued: 06 October 2005  
Renewal Date: 05 October 2010

Chief Executive

*MCERTS is operated on behalf of the Environment Agency by*

### **Sira Certification Service**

South Hill, Chislehurst, Kent, BR7 5EH, England  
Tel: 020 8467 2636 Fax: 020 8468 1841